

MRI Safety Information for Canada

1. English	MRI Safety Information for Canada	2
2. Français	Sécurité dans l'environnement de résonance magnétique pour le Canada	3

The Straumann® dental implants and dental implant abutments have not been evaluated for safety and compatibility in the MR environment. They have not been tested for heating, migration, or image artifact in the MR environment. The safety of Straumann® dental implants and dental implant abutments in the MR environment is unknown. Scanning a patient who has this device may result in patient injury.

La sécurité et la compatibilité des implants dentaires et des parties secondaires pour implants dentaires Straumann® n'ont pas été évaluées dans l'environnement de résonance magnétique. Ils n'ont pas été testés pour le risque potentiel de chauffage, de migration ou d'artéfacts d'imagerie dans un environnement de résonance magnétique. La sécurité des implants dentaires et parties secondaires pour implants dentaires Straumann® dans l'environnement de résonance magnétique n'est pas connue. L'examen d'un patient porteur de ce dispositif risque d'occasionner des lésions au patient.